

Listing and evaluation of the results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 24 October 2018



Adviser:

Dr. med. Christoph Niederau
MVZ Labor Dortmund Leopoldstraße GbR
Leopoldstraße 10
44147 Dortmund

Tel.: +49 231 860 27 0
Fax: +49 231 860 27 2810
Mail: christoph.niederau@labor-dortmund.de

296

Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Meets criteria	
Ascorbic acid (Vit. C)	61	mg/l	19.7	20.1	SV	12.9	27.3	-2.0%	+	
	62		9.70	9.89	SV	6.33	13.5	-1.9%	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	61	mg/l	19.7	1.00	9999	SD	ZY99
	62		9.70	1.00			



Ascorbic acid (Vit. C) (mg/l, N = 63)										
Collective	Sam- ple	Target value	Target range	Participants collective			Rate (%)			
				AVG	CV	Num.	Sam.	total		
all methods	61	20.1	12.9 - 27.3	20.1	10.9	63	95.2	92.1		
	62	9.89	6.33 - 13.5	9.89	11.3	63	92.1			

Rate of success: 92,1%

