

Listing and evaluation of the results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 23 October 2019



Adviser:

Dr. med. Christoph Niederau
MVZ Labor Dortmund Leopoldstraße GbR
Leopoldstraße 10
44147 Dortmund

Tel.: +49 231 860 27 0
Fax: +49 231 860 27 2810
Mail: christoph.niederau@labor-dortmund.de

296

Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Ascorbic acid (Vit. C)	61	mg/l	39.2	30.1	SV	19.3	40.9	30.2%	2,48	+	
	62		19.7	18.4	SV	11.8	25.0	7.1%	0,506	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	61	mg/l	39.2	1.00	9999	SD	ZY99
	62		19.7	1.00			

Individual summary of results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 23 October 2019

296

Vitamins 03 - Vitamin C

Ascorbic acid (Vit. C) (mg/l, N = 71)									
Collective	Sample	Target value	Target range	Participants collective			Rate (%)		
				AVG	CV	Num.	Sam.	total	
all methods	61	30.1	19.3 - 40.9	30.1	12.2	71	94.4	93.0	
	62	18.4	11.8 - 25.0	18.4	14.0	71	94.4		

Rate of success: 93%

