

Listing and evaluation of the results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 21 May 2021



Adviser:

Dr. med. Christoph Niederau
MVZ Labor Dortmund Leopoldstraße GbR
Leopoldstraße 10
44147 Dortmund

Tel.: +49 231 860 27 0
Fax: +49 231 860 27 2810
Mail: christoph.niederau@labor-dortmund.de

296

Vitamins 03 - Vitamin C

| Analyte | Sam- ple | Unit | Your value | Target value | TV- Type | Lower limit | Upper limit | Deviation | Z-Score | Meets criteria | |
|------------------------|-------------|------|---------------|-----------------|-------------|----------------|----------------|-----------|---------|-------------------|---|
| Ascorbic acid (Vit. C) | 31 | mg/l | 2.27 | 15.3 | SV | 9.79 | 20.8 | -85.2% | -9,05 | - | ○  |
| | 32 | | 3.15 | 27.8 | SV | 17.8 | 37.8 | -88.7% | -10,6 | - | ○  |

| Analyte | Sam- ple | Your unit | Stated value | Conversion factor | Method | Manu- facturer | Device |
|------------------------|-------------|-----------|--------------|----------------------|--------|-------------------|--------|
| Ascorbic acid (Vit. C) | 31 | mg/l | 2.27 | 1.00 | 9999 | SD | ZY99 |
| | 32 | | 3.15 | 1.00 | | | |



Individual summary of results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 21 May 2021

296

Vitamins 03 - Vitamin C

Ascorbic acid (Vit. C) (mg/l, N = 66)

| Collective | Sample | Target value | Target range | Participants collective | | | Rate (%) | |
|-------------|--------|--------------|--------------|-------------------------|------|------|----------|-------|
| | | | | AVG | CV | Num. | Sam. | total |
| all methods | 31 | 15.3 | 9.79 - 20.8 | 15.3 | 9.41 | 66 | 92.4 | 92.4 |
| | 32 | 27.8 | 17.8 - 37.8 | 27.8 | 8.38 | 66 | 92.4 | 92.4 |

Rate of success: 92,4%

