



Listing and evaluation of the results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 22 October 2021

Adviser:

Dr. med. Christoph Niederau

MVZ Labor Dortmund Leopoldstraße GmbH
Leopoldstraße 10
44147 Dortmund

Tel.: +49 231 860 27 0

Fax: +49 231 860 27 2810

Mail: christoph.niederau@labor-dortmund.de

296

Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Ascorbic acid (Vit. C)	61	mg/l	14.0	28.2	SV	18.0	38.4	-50.5%	-5,5	-	
	62		6.85	10.5	SV	6.72	14.3	-34.8%	-3,61	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	61	mg/l	14.0	1.00	9999	SD	ZY99
	62		6.85	1.00			



Individual summary of results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 22 October 2021

296

Vitamins 03 - Vitamin C

Ascorbic acid (Vit. C) (mg/l, N = 70)										
Collective	Sam-ple	Target value	Target range			Participants collective			Rate (%)	
						AVG	CV	Num.	Sam.	total
all methods	61	28.2	18.0	-	38.4	28.2	9.18	70	97.1	97.1
	62	10.5	6.72	-	14.3	10.5	9.62	70	98.6	

Rate of success: 97,1%

