

Listing and evaluation of the results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 20 May 2022

Adviser:

Dr. med. Christoph Niederau
MVZ Labor Dortmund Leopoldstraße GmbH
Leopoldstraße 10
44147 Dortmund

Tel.: +49 231 860 27 0
Fax: +49 231 860 27 2810
Mail: christoph.niederau@labor-dortmund.de

296

Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Ascorbic acid (Vit. C)	31	mg/l	14.2	22.1	SV	14.1	30.1	-35.9%	-3,32	+	
	32		4.77	4.48	SV	2.87	6.09	6.5%	0,497	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	31	mg/l	14.2	1.00	9999	SD	ZY99
	32		4.77	1.00			

Individual summary of results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 20 May 2022

296

Vitamins 03 - Vitamin C

Ascorbic acid (Vit. C) (mg/l, N = 84)										
Collective	Sam- ple	Target value	Target range			Participants collective			Rate (%)	
						AVG	CV	Num.	Sam.	total
all methods	31	22.1	14.1	-	30.1	22.1	10.8	84	94.0	86.9
	32	4.48	2.87	-	6.09	4.48	13.0	84	89.3	

Rate of success: 86,9%

