



Listing and evaluation of the results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 28 October 2022

Adviser:

Dr. med. Christoph Niederau

MVZ Labor Dortmund Leopoldstraße GmbH
Leopoldstraße 10
44147 Dortmund

Tel.: +49 231 860 27 0

Fax: +49 231 860 27 2810

Mail: christoph.niederau@labor-dortmund.de

296

Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Ascorbic acid (Vit. C)	61	mg/l	14.2	15.5	SV	9.92	21.1	-8.6%	-1	+	
	62		4.61	2.72	SV	1.74	3.70	69.5%	5,43	-	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	61	mg/l	14.2	1.00	9999	SD	ZY99
	62		4.61	1.00			



Individual summary of results

51396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 28 October 2022

296

Vitamins 03 - Vitamin C

Ascorbic acid (Vit. C) (mg/l, N = 86)										
Collective	Sample	Target value	Target range			Participants collective			Rate (%)	
						AVG	CV	Num.	Sam.	total
all methods	61	15.5	9.92	-	21.1	15.5	8.58	86	97.7	91.9
	62	2.72	1.74	-	3.70	2.72	12.8	86	93.0	

Rate of success: 91,9%

