



## Listing and evaluation of the results

51396: Dr. Endre Mikus  
LabMagister Training and Science Ltd.

Survey of 19 May 2023

Adviser:

Dr. med. Christoph Niederau

MVZ Labor Dortmund Leopoldstraße GmbH  
Leopoldstraße 10  
44147 Dortmund

Tel.: +49 231 860 27 0

Fax: +49 231 860 27 2810

Mail: christoph.niederau@labor-dortmund.de

296

### Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Ascorbic acid (Vit. C)	31	mg/l	21.6	21.6	SV	13.8	29.4	-0.2%	-0,025	+	
	32		10.5	10.2	SV	6.53	13.9	2.8%	0,39	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	31	mg/l	21.6	1.00	9999	SD	ZY99
	32		10.5	1.00			



## Individual summary of results

51396: Dr. Endre Mikus  
LabMagister Training and Science Ltd.

Survey of 19 May 2023

296

### Vitamins 03 - Vitamin C

Ascorbic acid (Vit. C) (mg/l, N = 86)										
Collective	Sam- ple	Target value	Target range			Participants collective			Rate (%)	
						AVG	CV	Num.	Sam.	total
all methods	31	21.6	13.8	-	29.4	21.6	7.45	86	95.3	95.3
	32	10.2	6.53	-	13.9	10.2	7.28	86	96.5	

Rate of success: 95,3%

