



Listing and evaluation of the results

10051396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 27 October 2023

Adviser:

Dr. med. Christoph Niederau

MVZ Labor Dortmund Leopoldstraße GmbH
Leopoldstraße 10
44147 Dortmund

Tel.: +49 231 860 27 0

Fax: +49 231 860 27 2810

Mail: christoph.niederau@labor-dortmund.de

296

Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Ascorbic acid (Vit. C)	61	mg/l	8.39	10.1	SV	6.46	13.7	-16.9%	-1,87	+	
	62		5.70	2.71	SV	1.73	3.69	110.3%	7,81	-	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	61	mg/l	8.39	1.00	9999	SD	ZY99
	62		5.70	1.00			

Additional information on the evaluation is available for download in the RV-Online system in the column "Report".

Individual summary of results

10051396: Dr. Endre Mikus
LabMagister Training and Science Ltd.

Survey of 27 October 2023

296

Vitamins 03 - Vitamin C

Ascorbic acid (Vit. C) (mg/l, N = 88)

Collective	Sam- ple	Target value	Target range	Participants collective			Rate (%)	
				AVG	CV	Num.	Sam.	total
all methods	61	10.1	6.46 - 13.7	10.1	9.04	88	94.3	89.8
	62	2.71	1.73 - 3.69	2.71	14.1	88	89.8	

Rate of success: 89,8%

