

Scheme: Sperm-hyaluronan assay

Distribution: Q4/4 - 2022/2023

Lab name: LabMagister

Lab number: 971

Reviewed by:


Date:

Action taken:


I Dr Mikus Endre on behalf of LabMagister have read and agreed to the terms and conditions outlined by EQASRM.

Sample code	HB413	HB414	HB415	HB416
Description	Semen sample I	Semen sample II	Semen sample III	Semen sample IV
Number of returns	3	3	3	3
Responses for normal maturity/physiological function	Yes 2/3 No 1/3	Yes 2/3 No 1/3	Yes 3/3 No 0/3	Yes 3/3 No 0/3
Your result (% sperm bound to slide)	73.1	57	88	89
Does this indicate normal maturity/physiological function?	No	No	Yes	Yes
Comment	Acceptable	Acceptable	Acceptable	Acceptable
Overall statistic				
Mean	81.4	71.7	89	89.7
SD	7.3	13.3	1.7	1.2
Range	[73.1-87]	[57-83]	[88-91]	[89-91]
Your bias	-10.2%	-20.5%	-1.1%	-0.8%
Points to note				


HB413




HB414




HB415



HB416



Bias



You can alter the date range of results by grabbing and sliding the ends of this section.

