



## Listing and evaluation of the results

10051396: Dr. Endre Mikus  
LabMagister Training and Science Ltd.

Survey of 10 May 2024

Adviser:

Dr. med. Christoph Niederau

MVZ Labor Dortmund Leopoldstraße GmbH  
Leopoldstraße 10  
44147 Dortmund

Tel.: +49 231 860 27 0

Fax: +49 231 860 27 2810

Mail: christoph.niederau@labor-dortmund.de

296

### Vitamins 03 - Vitamin C

Analyte	Sam- ple	Unit	Your value	Target value	TV- Type	Lower limit	Upper limit	Deviation	Z-Score	Meets criteria	
Ascorbic acid (Vit. C)	31	mg/l	20.6	26.6	SV	17.0	36.2	-22.7%	-2,43	+	
	32		2.80	2.69	SV	1.72	3.66	4.1%	0,276	+	

Analyte	Sam- ple	Your unit	Stated value	Conversion factor	Method	Manu- facturer	Device
Ascorbic acid (Vit. C)	31	mg/l	20.6	1.00	9999	SD	ZY99
	32		2.80	1.00			

Additional information on the evaluation is available for download in the RV-Online system in the column "Report".



## Individual summary of results

10051396: Dr. Endre Mikus  
LabMagister Training and Science Ltd.

Survey of 10 May 2024

296

### Vitamins 03 - Vitamin C

Ascorbic acid (Vit. C) (mg/l, N = 89)										
Collective	Sample	Target value	Target range			Participants collective			Rate (%)	
						AVG	CV	Num.	Sam.	total
all methods	31	26.6	17.0	-	36.2	26.6	9.32	89	89.9	86.5
	32	2.69	1.72	-	3.66	2.69	14.8	89	89.9	

Rate of success: 86,5%

